

MxP[®] Quant 500 kit requirements

This document specifies additional lab equipment and chemicals that are required but <u>not</u> included with the kit. Before starting with the kit, ensure that all required items are available.

LC-MS system (1|2)

Agilent	SCIEX		
Mass spectrometer			
6495D, 6495C or 6495B Triple Quadrupole LC/MS System	7500, 6500+, 6500, 5500+ or 5500 series (QTRAP or Triple Quad) with TurboV™ ion source or Optiflow Pro (high flow)		
LC and autosampler			
 Agilent 1290 Infinity I/II UHPLC system with column oven. NanoLC and microLC are not supported! 800 bar injection valve (Agilent part no. 5067-6698) replaces standard 1300 bar valve 1290 Infinity II inline filters, 0.3 µm (2 pcs., before and after injection valve, Agilent part no. 5067-6189) 96-well plate autosampler with temperature control (10 °C) Injection volume range: 2 - 20 µL 	 Standard flow HPLC or UHPLC system with column oven. NanoLC and microLC are not supported! 96-well plate autosampler with temperature control (10 °C) Injection volume range: 5 - 20 µL Special sample rack for deep-well plates required for the following autosamplers: Shimadzu: part no. 228-37546-92 (Deep-well MTP Rack, DWP-96) UltiMate 3000: Thermo part no. 6820.4083 (30 - 36 mm height) 		
Tubing			

- Recommended between injector and column system: steel capillary
- Recommended between column system and mass spectrometer: red PEEK Tubing 1/16" OD x 0.005" ID



LC-MS system (2|2)

Thermo	Waters			
Mass spectrometer				
TSQ Altis or Altis Plus with OptaMax NG H-ESI source and argon collision gas	Xevo TQ Absolute, TQ-XS or TQ-S with ESI source			
LC and autosampler				
 Thermo Vanquish UHPLC system with default configuration: 35 µL mixer set, Vanquish H Pump (6044.5018) 25 µL sample loop (6850.1911) Vanquish active pre-heater, MP35N, 0.1 x 380 mm (6732.0110) Vanquish post-column cooler 1 µL, 0.1 x 240 mm (6732.0510) SII for Xcalibur 1.6 or later Vanquish firmware 2.01 or later 	 Waters ACQUITY UPLC[®] system with sample manager, solvent manager, and column oven: Classic (loop-based) I-Class (FTN) H-Class (FTN) Injection volume range: 2 - 20 μL FTN samplers require extension loop of 50 μL or needle size of 30 μL! 			
Tubing				
 Viper MP35N, 0.1 x 350 mm (6042.2340) – Divert valve to MS Viper PEEK, 0.13 x 150 mm (6041.5616) – Ground union to spray needle Recommended union (U-402) for FIA 	 Recommended between injector and column system: steel capillary Recommended between column system and mass spectrometer: red PEEK Tubing 1/16" OD x 0.005" ID 			

LC-MS instrument status

Condition

- Serviced, calibrated and tuned instruments according to manufacturer's recommendations on a regular basis
- LC-MS system must be free of ion-pair reagent

Analytical column

HPLC and UHPLC		
Analytical column	MxP [®] Quant 500 (XL) kit column system: biocrates part number 21117.1 (UHPLC), 21117.2 (HPLC) (HPLC/UHPLC column + pre-column mixer + connector, factory assembled)	



Laboratory equipment

Laboratory equipment (may also be available from your local kit distributor)			
	Nitrogen evaporator for 96-well plates*	Pressure manifold for 96-well plates*	
Nitrogen evaporator <u>or</u> pressure manifold - Requires nitrogen supply - Must be in a fume hood - No heater required	 Examples: Porvair blowdown evaporator MiniVap® Techne (FSC496D) Sample concentrator from BenchTop Lab Systems (96 samples, BT1604) VLM evaporators Organomation MICROVAP microplate evaporator Biotage TurboVap® 96 	 Examples: Porvair UltraPPM LITE Waters positive pressure-96 processor (186006961) Biotage[®] PRESSURE+ 96 manifold (PPM-96) CEREX[®] system 96 processor (288-0001) Agilent positive pressure manifold 96 processor (PPM-96) TECAN Resolvex[®] M10 96 	
Centrifuge	Must be able to centrifuge 96-well plates of 5 cm height at 500 x g	Not required when a pressure manifold is used.	
Shaker	Any model with adjustable speed (450 – 1200 rpm) including a tray for plates and vials. Recommended: Eppendorf MixMate® or ThermoMixer®		
Pipettes	 Repeater, e.g. Eppendorf Multipette[®] E3 (with 2.5 and 10 mL tips) or similar electronic model Single channel: volume range 10 μL – 1000 μL 8-Channel: 10 μL – 150 μL 		
Vortexer	Any model		
Balance	Accuracy < 1 mg		
Solvent bottles	50 – 1000 mL		

* See document "Technical guide-Nitrogen evaporators and pressure manifolds" for details.

Automation (optional)

Waters Andrew+ pipetting robot		
Andrew+ system components and protocol specifications:		
https://dev-onelab.andrewalliance.com/library/mxp-quant-500-kit-automated-sample-prep- method-bQ0eLAMI		



Urine extension (if applicable)

O mly fai				samples
$(J \Pi V I \Omega)$	I USEIS A	nav/ma	urine	Samples
			Grine	oumpree

biocrates part number 21933.1

Solvents and chemicals

Solvents and chemicals	Purity
Ethanol, methanol, water, acetonitrile, isopropanol	LC-MS grade (ethanol: HPLC grade sufficient)
Formic acid	LC-MS grade (e.g. Honeywell Fluka [™] 56302-50ML) Fresh! New bottle or opened within past 6 months
Phenyl isothiocyanate (PITC)	99%, for protein sequencing (e.g. Sigma-Aldrich 317861) Fresh! New bottle or opened within past 6 months
Pyridine - or -	Purity >99% (recommended Acros Organics 131780500, 50mL) Fresh! New bottle or opened within past 6 months
Triethylamine (Urine only)	Purity ≥99.5% (e.g. Sigma-Aldrich 90335-100ML)
Ammonium acetate	LC-MS grade
Phosphate buffered saline (PBS)	p.a. grade (e.g. Sigma-Aldrich P4417)

Software

Agilent	SCIEX	Thermo	Waters	
MS software				
MassHunter version 10.1 or 12	Analyst® version 1.7 or later SCIEX OS version 3 or later	XCalibur™ 4.5 or later SII 1.6 or later Chromeleon™ CDS 7.3 or later (please first contact biocrates support)	MassLynx® version 4.1 or later	
biocrates WebIDQ workflow manager				

Please get one of the three subscription versions through our webshop or your local distributor:

- WebIDQ cloud subscription (10 GB storage space) biocrates part number 21629
- WebIDQ cloud+ subscription (50 GB storage space) biocrates part number 21926
- WebIDQ on-premises subscription biocrates part number 21636

For detailed system requirements please visit:

https://shop.biocrates.com/faq/What-are-the-system-requirements-for-WebIDQ