

# MxP® Quant HR Xpress kit requirements

This document specifies additional lab equipment and chemicals that are required but <u>not</u> included with the kit. Before starting with the kit, ensure that all required items are available.

# LC-MS system

| Thermo Scientific  |   |  |
|--|---|--|
| Q Exactive with <u>HESI-II</u> ion source  | Orbitrap Exploris   |  |
| <ul> <li>Q Exactive™</li> <li>Q Exactive™ Focus</li> <li>Q Exactive™ Plus</li> <li>Q Exactive™ HF</li> <li>Q Exactive™ HF-X</li> </ul> | <ul><li>Orbitrap Exploris 120</li><li>Orbitrap Exploris 240</li><li>Orbitrap Exploris 480</li></ul> |  |

### LC and autosampler

- Standard flow UHPLC system with column oven. NanoLC and microLC are not supported!
- 96-well plate autosampler with temperature control (10 °C)
- Injection volume range: 5 20 μL
- Special sample rack for deep-well plates required for the following autosamplers:
  - UltiMate 3000: Thermo part no. 6820.4083 (30 36 mm height)

### **Tubing**

- Required between injector and column system:
   Thermo Scientific™ Viper™ Capillary 0.13 x 350 mm (6040.2335)
- Recommended between column and mass spectrometer:
   Red PEEK Tubing 1/16" OD x 0.005" ID

### LC-MS instrument status

### Condition

- Serviced, calibrated and tuned instruments according to manufacturer's recommendations on a regular basis
- LC-MS system must be free of ion-pair reagent

## Analytical column

|                   | UHPLC   |
|-------------------|---|
| Analytical column | MxP® Quant HR Xpress kit UHPLC column: biocrates product number 9120052121254 |



# Solvents and chemicals

| Solvents and chemicals                              | Purity  |
|---|---|
| Acetonitrile, ethanol, isopropanol, methanol, water | LC-MS grade (ethanol: HPLC grade sufficient)  |
| Formic acid   | LC-MS grade (e.g. Honeywell Fluka™ 56302-50ML) Fresh! New bottle or opened within past 6 months           |
| Phenyl isothiocyanate (PITC)                        | 99%, for protein sequencing (e.g. Sigma-Aldrich 317861) Fresh! New bottle or opened within past 6 months  |
| Pyridine  | Purity >99% (recommended Acros Organics 131780500, 50mL) Fresh! New bottle or opened within past 6 months |
| Ammonium acetate                                    | LC-MS grade   |
| Phosphate buffered saline (PBS)                     | p.a. grade (e.g. Sigma-Aldrich P4417)   |

# Laboratory equipment

| Laboratory equipment (may also be available from your local kit distributor)                                  |  |   |  |
|---|--|---|--|
|   | Nitrogen evaporator<br>for 96-well plates  | Pressure manifold<br>for 96-well plates   |  |
| <ul><li>Nitrogen evaporator*</li><li>or</li><li>pressure manifold*</li><li>Requires nitrogen supply</li></ul> | Examples:  - Porvair blowdown evaporator MiniVap®  - Techne (FSC496D)  - Sample concentrator from BenchTop Lab Systems   | Examples:  - Porvair UltraPPM LITE  - Waters positive pressure-96 processor (186006961)  - Biotage® PRESSURE+ 96 manifold (PPM-96)                          |  |
| <ul><li>Must be in a fume<br/>hood</li><li>No heater required</li></ul>                                       | (96 samples, BT1604)  - VLM evaporators  - Organomation MICROVAP microplate evaporator  - Biotage TurboVap® 96   | <ul> <li>CEREX® system 96 processor (288-0001)</li> <li>Agilent positive pressure manifold 96 processor (PPM-96)</li> <li>TECAN Resolvex® M10 96</li> </ul> |  |
| Centrifuge  | Must be able to centrifuge 96-well plates of 5 cm height at 500 x g  | Not required when a pressure manifold is used.  |  |
| Shaker  | Any model with adjustable speed (450 – 1200 rpm) including a tray for plates and vials. Recommended: Eppendorf MixMate® or ThermoMixer®  |   |  |
| Pipettes  | <ul> <li>Repeater, e.g. Eppendorf Multipette® E3         <ul> <li>(with 1.0, 2.5 and 10 mL tips) or similar electronic model</li> <li>Single channel: volume range 10 μL – 1000 μL</li> <li>8-Channel: 150 μL</li> </ul> </li> </ul> |   |  |
| Vortexer  | Any model  |   |  |
| Balance   | Accuracy < 1 mg  |   |  |
| Solvent bottles   | 50 – 1000 mL   |   |  |

 $<sup>\</sup>hbox{$^*$ See document "Technical guide-Nitrogen evaporators and pressure manifolds" for details.}$ 



### Software

### Thermo Scientific

### MS software

Xcalibur™ software 4.0 or later

## biocrates WebIDQ workflow manager

Please get one of the two subscription versions through our webshop or your local distributor:

- WebIDQ cloud+ subscription (50 GB storage space) biocrates part number 21926
- WebIDQ on-premises subscription biocrates part number 21636

For detailed system requirements please visit:

https://shop.biocrates.com/faq/What-are-the-system-requirements-for-WebIDQ